




HKN-90

INFANT RADIANT WARMER



Brief Introduction

- Three control Modes: Pre-warm Mode, Manual Mode and Baby Mode controlled by micro-computer;
- Set Temperature and Baby Temperature can be displayed separately;
- Horizontal angle of warmer module and inclination of the bassinet can be adjusted;
- The panels surrounded bassinet can be disassembled;
- APGAR timer;
- Multiple failure alarm indications;
- RS232 connector for data output;
- Disposable skin temperature sensor is optional.

Standard Configuration

Main body (including the warmer module, temperature controller, bassinet, stand), I.V. Pole, tray.

Specification

Power requirement: AC220V-230V/50Hz or AC110-120V/50-60Hz, 600VA
Control Mode: Pre-warm Mode, Manual Mode and Baby Mode control by micro-computer
Baby Mode temperature control range: 34.5-37.5°C
Skin temperature display range: 5°C-65°C
Accuracy of skin temperature sensor: ±0.2°C
Temperature uniformity of mattress: ≤2°C
Angle of heater head: 0°, 30°, 45° for two ways
Inclination of the bassinet: five tilt position
APGAR timer: Audible and visible alarm when the device runs to 50" ~ 1' , 4' 50" ~ 5' , 9' 50" ~ 10'
Failure alarm: over temperature alarm, deviation alarm, sensor failure alarm, power failure alarm, setting alarm, checking alarm and so on.

Environmental Requirement

Operating range: 18°C-30°C
Ambient air velocity rate: <0.3m/s

Product Packaging

Package: The whole unit is packed into one carton
Dimension: 1110mm x 625mm x 1195mm
Gross Weight: 60kg



Stock code [300314]



Ningbo David Medical Device Co., Ltd.

Sales center: #100 Jinghua Road, High-tech Industrial Development Zone, Ningbo, China
Tel: 0086-574-8780 0008 8780 0007 /Fax: 0086-574-8780 3714
E-mail: sales@nbdavid.com /Http: //www.chinadavid.cn /P.C.: 315040